



Product Service

Attestation of Conformity

No. M8A 117743 0001 Rev. 00

Holder of Certificate: Maabarot Metal Works Ltd

12 Bazelet St,
4486200 Tzur Yigal
ISRAEL

**Product: Electrical Machinery
(Medical Waste Disinfection System)**

Model(s): Envomed 80

Parameters:

Rated Input Voltage	: 380-400VAC, 3P + N + PE
Rated Frequency	: 50/60 Hz
Rated Input Current	: 32 A
Protection Class	: I

Tested according to: EN ISO 12100:2010
EN 60204-1:2018

This Attestation of Conformity is issued on a voluntary basis according to Council Directive 2006/42/EC relating to machinery. It confirms that the listed equipment (not Annex IV equipment) complies with the principal protection requirements of the directive. It refers only to the sample submitted to TÜV SÜD Product Service GmbH for testing and certification. For details see: www.tuvsud.com/ps-cert

Test report no.: 7191287912-EEC22-TZ

Date, 2022-07-28

(QI THOMAS LUO)